

AMDA-Nepal

Siddhartha Children And Women Hospital

Butwal-7, Rupandehi

Technical Specification hospital beds Double folding with side rallying

S.N.	Purchaser's Specifications	Bidder's Compliance (Yes/No)	Reference Page No.	Remarks
	Hospital Beds Adults			
	Manufacturer:			
	Brand:			
	Type / Model:			
	Country of Origin:			
1	Technical specification			
2	Double folding bed with a side railing typically features a compact, space-saving metal frame, with epoxy coated powder a folding mechanism, and a protective safety guardrail.			
3	Features both head and (knee-rest elevation) Standard single size, roughly 200 cm × 80 cm to 90 cm (or around 6 feet × 3 feet)			
4	Supports between 150 kg to 200kg for standard adult.			
5	Side Railing Height: More than 60 cm (24inches) from bed height to prevent rolls or falls.			
6	Rail Operation: Foldable or drop-down guardrail operated via a spring-loaded or push-button mechanism			
7	Back-rest Adjustment: 0-75°(±10°). Knee-rest Adjustment: 0-35°(±10°).			
8	Side Rail Locating Bar: Ensures secure positioning of side rails.			
9	Casters with Brakes: Double-sided casters equipped with bark for stability.			
10	GUARD RAIL LIFTING The height of the guard rails could be adjusted in two positions. It adopts aluminum alloy material inlaid with transparent acrylic panels, convenient to observe the children's status			
11	Head and foot rest can be manually raised more by the ratchet mechanism.			
12	Bumper wheels on each corner: 4pcs.			
13	With IV Pole: 1 set			

14	Optional Accessories: Mattress, patient lifter, over bed table.			
15	Detail Instruction Head and Foot Board Detachable head and foot board, once step quick removable Epoxy Steel Painted Safe Side Rail Full length side rail. extremely safety Side Rail Castor Built-in controller.			
16	Mattress: Often includes an integrated foam mattress ranging from 4 to 6 inches thick (though some frames are sold separately.			
17	Standards and Safety Requirements			
18	Must submit ISO13485. Certificate			
19	Warranty			
20	Comprehensive warranty for 1 years in all parts (wheel, crank) after acceptance.			
21	Maintenance Service During Warranty Period During warranty period supplier must ensure preventive maintenance & corrective/breakdown maintenance whenever required.			
22	Documentation User (Operating) manual in English Certificate of calibration and inspection from factory. Must submit valid authorization letter			