

AMDA-Nepal

Siddhartha Children And Women Hospital
Butwal-7, Rupandehi

Technical Specification of TVS probe for Ultrasound machine

S.N.	Purchaser's Specifications	Bidder's Compliance (Yes/No)	Reference Page No.	Remarks
	Samsung: HS40			
	Manufacturer:			
	Brand:			
	Type / Model:			
	Country of Origin:			
1	Transvaginal (TVS) probe is high-frequency, endocavitary transducer designed for detailed imaging of the female pelvic organs. Application on OB,GYN, Urology			
1.1	Probe frequency : 4-9 MHz			
1.2	Radius of curvature :10mm			
1.3	Field of view: 145 degree or better performance.			
1.4	Warranty: One year warranty.			
1.5	The bidder must submit a valid authorization letter from the manufacture company.			

